

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/05/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:10
INTERNAL STANDARD	VERIFIED	11:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:11
BLANK TEST	0.00	11:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:12
BLANK TEST	0.00	11:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:13
BLANK TEST	0.00	11:14

