

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 09/01/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:38
INTERNAL STANDARD	VERIFIED	09:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:39
BLANK TEST	0.00	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:40
BLANK TEST	0.00	09:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:41
BLANK TEST	0.00	09:42

