

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 09/30/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:39
INTERNAL STANDARD	VERIFIED	09:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:39
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:40
BLANK TEST	0.00	09:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:41
BLANK TEST	0.00	09:42

25-44394

