

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/07/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:44
INTERNAL STANDARD	VERIFIED	10:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:45
BLANK TEST	0.00	10:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:47
BLANK TEST	0.00	10:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:48
BLANK TEST	0.00	10:49

