

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/14/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:36
INTERNAL STANDARD	VERIFIED	11:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:37
BLANK TEST	0.00	11:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:38
BLANK TEST	0.00	11:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:39
BLANK TEST	0.00	11:40

