

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/29/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:40
INTERNAL STANDARD	VERIFIED	09:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:40
BLANK TEST	0.00	09:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:41
BLANK TEST	0.00	09:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:43
BLANK TEST	0.00	09:43

25-4888

