

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/07/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:53
INTERNAL STANDARD	VERIFIED	14:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:54
BLANK TEST	0.00	14:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:55
BLANK TEST	0.00	14:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:56
BLANK TEST	0.00	14:57

2MS
75-1092

