

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/21/2025

*RMS
25-3130*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:20
INTERNAL STANDARD	VERIFIED	13:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	13:20
BLANK TEST	0.00	13:21
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	13:22
BLANK TEST	0.00	13:22
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	13:23
BLANK TEST	0.00	13:24

