

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/28/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:27
INTERNAL STANDARD	VERIFIED	10:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:27
BLANK TEST	0.00	10:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:28
BLANK TEST	0.00	10:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:30
BLANK TEST	0.00	10:30

RMJ
25-4395

