

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 02/03/2025

*RMS
25-5217*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:51
INTERNAL STANDARD	VERIFIED	08:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:51
BLANK TEST	0.00	08:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:52
BLANK TEST	0.00	08:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:53
BLANK TEST	0.00	08:54

