

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 02/04/2025

*RMS*  
*25-5387*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:28
INTERNAL STANDARD	VERIFIED	09:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:29
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:31
BLANK TEST	0.00	09:32

