

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 02/11/2025

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 24410

*DM>*  
*25-6741*

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:00
INTERNAL STANDARD	VERIFIED	10:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:01
BLANK TEST	0.00	10:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:02
BLANK TEST	0.00	10:03

