

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 02/18/2025

RMS
25-7155

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:48
INTERNAL STANDARD	VERIFIED	11:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:48
BLANK TEST	0.00	11:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:49
BLANK TEST	0.00	11:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:50
BLANK TEST	0.00	11:51

