

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 03/03/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 1278

RMS
25-4393

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:35
INTERNAL STANDARD	VERIFIED	11:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:35
BLANK TEST	0.00	11:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:36
BLANK TEST	0.00	11:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:37
BLANK TEST	0.00	11:38

