

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

*RMS*  
*25-9875*

Date: 03/04/2025

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:05
INTERNAL STANDARD	VERIFIED	11:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:06
BLANK TEST	0.00	11:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:07
BLANK TEST	0.00	11:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:08
BLANK TEST	0.00	11:09

