

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 03/18/2025

*2ms*  
*25-11879*

OPERATOR NAME:  
MICHAEL J CLIUVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:38
INTERNAL STANDARD	VERIFIED	12:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:39
BLANK TEST	0.00	12:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:40
BLANK TEST	0.00	12:41

