

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 04/08/2025

RM)
25-15/27

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:51
INTERNAL STANDARD	VERIFIED	09:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:51
BLANK TEST	0.00	09:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:52
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:53
BLANK TEST	0.00	09:54

