

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 06/18/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:28
INTERNAL STANDARD	VERIFIED	08:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:29
BLANK TEST	0.00	08:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:30
BLANK TEST	0.00	08:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:31
BLANK TEST	0.00	08:32

RM's
25-27582

