

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 07/08/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:22
INTERNAL STANDARD	VERIFIED	08:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:22
BLANK TEST	0.00	08:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:25
BLANK TEST	0.00	08:25

RMS
25-30610

