

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 07/15/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:30
INTERNAL STANDARD	VERIFIED	10:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:30
BLANK TEST	0.00	10:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:32
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:33

RMS
25-32000

