

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 07/22/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25110

RMS
25-33511

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:54
INTERNAL STANDARD	VERIFIED	07:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:55
BLANK TEST	0.00	07:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:56
BLANK TEST	0.00	07:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58

