

# SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 08/05/2025

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 25170

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 09:59 |
| INTERNAL STANDARD       | VERIFIED | 09:59 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 10:00 |
| BLANK TEST              | 0.00     | 10:00 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 10:01 |
| BLANK TEST              | 0.00     | 10:02 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 10:02 |
| BLANK TEST              | 0.00     | 10:03 |

