

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 08/26/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:49
INTERNAL STANDARD	VERIFIED	09:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:49
BLANK TEST	0.00	09:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:50
BLANK TEST	0.00	09:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:51
BLANK TEST	0.00	09:52

