

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 09/02/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:21
INTERNAL STANDARD	VERIFIED	12:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:22
BLANK TEST	0.00	12:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:23
BLANK TEST	0.00	12:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:24
BLANK TEST	0.00	12:25

