

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 09/16/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:24
INTERNAL STANDARD	VERIFIED	12:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:25
BLANK TEST	0.00	12:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:26
BLANK TEST	0.00	12:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:27
BLANK TEST	0.00	12:28

