

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 09/19/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:49
INTERNAL STANDARD	VERIFIED	15:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:49
BLANK TEST	0.00	15:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:50
BLANK TEST	0.00	15:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:51
BLANK TEST	0.00	15:52

2ms
25-43062

