

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 11/19/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:10
INTERNAL STANDARD	VERIFIED	09:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:10
BLANK TEST	0.00	09:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:11
BLANK TEST	0.00	09:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:12
BLANK TEST	0.00	09:13

