

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/27/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26270

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:50
INTERNAL STANDARD	VERIFIED	12:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:50
BLANK TEST	0.00	12:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:51
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:52
BLANK TEST	0.00	12:53

26-43414

