

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 09/22/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:56
INTERNAL STANDARD	VERIFIED	08:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:56
BLANK TEST	0.00	08:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:57
BLANK TEST	0.00	08:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:58
BLANK TEST	0.00	08:59

26-39291

