

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 09/29/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:18
INTERNAL STANDARD	VERIFIED	09:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:18
BLANK TEST	0.00	09:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:19
BLANK TEST	0.00	09:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:20
BLANK TEST	0.00	09:21

26-39569

