

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/02/2026

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:08
INTERNAL STANDARD	VERIFIED	15:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:08
BLANK TEST	0.00	15:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:09
BLANK TEST	0.00	15:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:10
BLANK TEST	0.00	15:11

RMS
20-210

