

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308




Date: 05/01/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:46
INTERNAL STANDARD	VERIFIED	09:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:47
BLANK TEST	0.00	09:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:48
BLANK TEST	0.00	09:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:49
BLANK TEST	0.00	09:50

26-16522



SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 05/01/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:50
INTERNAL STANDARD	VERIFIED	09:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:51
BLANK TEST	0.00	09:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:52
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:53
BLANK TEST	0.00	09:54

26-~~1~~6522

