

SOLUTION CHANGE

-----  
UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308  
-----




Date: 07/01/2026

OPERATOR NAME:  
JOSHUA R HARRINGTON  
Permit #: 38303  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:23
INTERNAL STANDARD	VERIFIED	12:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:23
BLANK TEST	0.00	12:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:24
BLANK TEST	0.00	12:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:25
BLANK TEST	0.00	12:26

26-25876



SOLUTION CHANGE

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 07/01/2026

OPERATOR NAME:  
JOSHUA R HARRINGTON  
Permit #: 38303  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:23
INTERNAL STANDARD	VERIFIED	12:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:23
BLANK TEST	0.00	12:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:24
BLANK TEST	0.00	12:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:25
BLANK TEST	0.00	12:26

26-25876

