

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 10/20/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:57
INTERNAL STANDARD	VERIFIED	09:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:57
BLANK TEST	0.00	09:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:00

26-42605

