

# SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 10/27/2026

OPERATOR NAME:  
JOSHUA R HARRINGTON  
Permit #: 38303  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 26270

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 12:50 |
| INTERNAL STANDARD       | VERIFIED | 12:50 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 12:50 |
| BLANK TEST              | 0.00     | 12:51 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 12:51 |
| BLANK TEST              | 0.00     | 12:52 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 12:53 |
| BLANK TEST              | 0.00     | 12:53 |

26-43414

