

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 09/16/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:43
INTERNAL STANDARD	VERIFIED	12:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:43
BLANK TEST	0.00	12:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:44
BLANK TEST	0.00	12:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:45
BLANK TEST	0.00	12:46

26-37577

