

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 01/02/2026

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:04
INTERNAL STANDARD	VERIFIED	15:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:05
BLANK TEST	0.00	15:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:06
BLANK TEST	0.00	15:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:07
BLANK TEST	0.00	15:08

2MS
26-210

