

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 01/06/2026

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:31
INTERNAL STANDARD	VERIFIED	10:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:31
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:32
BLANK TEST	0.00	10:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:34

