

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308


Date: 02/02/2026

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:35
INTERNAL STANDARD	VERIFIED	10:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:35
BLANK TEST	0.00	10:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:37
BLANK TEST	0.00	10:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:38
BLANK TEST	0.00	10:39

26-4191



SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 02/02/2026

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:59
INTERNAL STANDARD	VERIFIED	09:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:01
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:02
BLANK TEST	0.00	10:02

26-4191

